



Certificate of Calibration

Certificate Number : 12345678

Instrument Details :

Manufacturer : MEET Electronics Limited

Description : Pin + Pinless Moisture Meter

Model Number : MS-98TGSE

Serial Number : 000000

ID Number : N/A

Calibration Date :

Intervals : 12 months

Environmental Details :

Temperature : 25°C ±5°C

Relative Humidity : 60% ±15%

MEET certifies that the instrument listed above meets the specifications of the manufacturer as the completion of its calibration. This certificate is not to be reproduced other than in full, except with prior written approval of MEET Electronics Limited. All the calibration standards used have an accuracy or better, unless otherwise stated.

Approved By :

Calibration Condition : Pass



MEET Electronics Ltd.
QC Manager



List of used Test Equipment :

Description	Manufacturer and model	Serial No.	Calibration Lab
Test Block	MEET, MS-CTBP1	MS-CTBP11	Traceable
Test Block	MEET, MS-CTBPL1	MS-CTBPL11	Traceable

1. Calibration Test Block (MS-CTBP1) for Pin type :

Wood Mode (Pin type & Remote Pin type)

	Min.	Reading	Max.
Test Block 1	13.0	13.9	15.0
Test Block 2	23.0	24.0	25.0

Building Material (Pin type & Remote Pin type)

	Min.	Reading	Max.
Test Block 1	21.0	21.9	23.0
Test Block 2	33.0	33.9	35.0

2. Calibration Test Block (MS-CTBPL1) for Pinless type :

Wood Mode (Pinless type)

	Min.	Reading	Max.
Test Block (High)	90	101	120
Test Block (Low)	145	160	175

Building Material (Pinless type)

	Min.	Reading	Max.
Test Block (High)	130	143	160
Test Block (Low)	175	188	205

3. Self Calibration Check Point : 18.2% \pm 1% reading

Wood Mode (Pin type) : 18.2%