



Certificate of Calibration

Certificate Number : 12345678

Instrument Details :

Manufacturer : MEET Electronics Limited

Description : Pin + Pinless Moisture Meter

Model Number : MS-W98U4

Serial Number :123456

ID Number : N/A

Calibration Due :

Intervals : 12 months

Environmental Details :

Temperature : 25°C ±5°C

Relative Humidity : 60% ±15%

MEET certifies that the instrument listed above meets the specifications of the manufacturer as the completion of its calibration. This certificate is not to be reproduced other than in full, except with prior written approval of MEET Electronics Limited. All the calibration standards used have an accuracy or better, unless otherwise stated.

Approved By :

Calibration Condition : Pass

MEET Electronics Ltd.
QC Manager



List of used Test Equipment :

Description	Manufacturer and model	Serial No.	Calibration Lab
Test Block	MEET, MS-CTBP1	MS-CTBP11	Traceable
Test Block	MEET, MS-CTBPL1	MS-CTBPL11	Traceable

1. Calibration Test Block (MS-CTBP1) for Pin type :

Pin Probe mode (%WME)

	Min.	Reading	Max.
Test Block 1	13.0		15.0
Test Block 2	23.0		25.0

2. Calibration Test Block (MS-CTBPL1) for Pinless type :

Pinless mode: Relative reading (REL)

	Min.	Reading	Max.
Test Block (High)	2.0		6.0
Test Block (Low)	6.0		11.0

3. Self Calibration Check Point : 18.2% ±1% reading

Pin Probe mode (%WME) : %